



CVIP, Inc.
801 Broad Street
Emmaus, PA 18049

Ph: 610-967-1525
Fax: 610-967-0303

PRESSURE TEST REPORT

(PNEUMATIC) / (HYDROSTATIC)

Job Number: 2594 Date: 11-19-05

Description: BNL PULSED MAGNET CRYostat

Slowly Increase Pressure as Follows and Hold for 10 Minutes.

<u>50</u> psia	@	<u>7:30</u>	Complete @	<u>7:40</u>
<u>100</u> "	@	<u>7:40</u>	Complete @	<u>7:50</u>
<u>150</u> "	@	<u>7:50</u>	Complete @	<u>8:00</u>
<u>200</u> "	@	<u>8:00</u>	Complete @	<u>8:10</u>
<u>225</u> "	@	<u>8:10</u>	Complete @	<u>8:20</u>

Slowly Increase Pressure to the Following and Hold for 10 Minutes.

243 psia @ 8:20 Complete @ 8:30

Slowly Decrease Pressure to the Following and Hold for - Minutes.

- @ - Complete @ -

Test Results: OK

Test Performed By: H. Muso

Gage #: PI-14

Cert #: 053104 - PI-14

Cal Date: 5/31/05

Due Date: 5/31/06

Q.C. Witness: Mark Muso Date: 11-21-05

AI Witness: _____ Date: _____

Customer Witness: _____ Date: _____